



# **MicroTester G2**

**Micro-scale Tension-Compression Test System**

***User Manual***  
***version 4.8***



## *Mechanical measurement and analysis of biomedical materials*

CellScale provides scientific and medical researchers with turn-key systems for measuring the mechanical properties of biomaterials.

Our focus is soft tissue – including skin, ligaments, blood vessels, heart valves, sclera, membranes, and scaffolds. We provide user-friendly software, easy-to-use patented attachment systems, and effective data analysis tools.

Our foundation was laid at one of the world's leading research institutions – the University of Waterloo. We understand research and aim to provide effective solutions at a reasonable price.

Explore our web site [www.cellscale.com](http://www.cellscale.com) or contact us to learn more about our measurement systems.

© 2022 CellScale. All rights reserved. This material may not be reproduced, displayed, modified or distributed without the express prior written permission of the copyright holder. For permission, contact CellScale Biomaterials Testing at [info@cellscale.com](mailto:info@cellscale.com).



Sales and Support  
CellScale Biomaterials Testing  
564 Weber St N, Unit 11, Waterloo, Ontario  
Canada N2L 5C6  
Phone: 519.342.6870

# Table of Contents

1. General Information .....	1
System Assembly .....	1
Connections to Supply .....	1
Safety Warnings .....	1
Operating Controls.....	1
General Maintenance .....	1
Approvals and Certification .....	2
2. Parts .....	3
Tools.....	3
Basic Components.....	3
3. Testing Terminology .....	4
Multiphase Test Cycles.....	4
Phases, Cycles, and Test Sequences.....	5
Test Phases: The Smallest Unit of Testing .....	6
Control Modes .....	7
Control Functions.....	7
4. System Overview .....	8
Microbeam.....	8
Microbeam Actuators.....	10
Imaging System .....	10
Fluid Bath .....	10
Software Overview.....	12
Output Files and Data Structures .....	14
5. Setting Up & Starting a Test.....	15
Overview .....	15
Step 1: Start a New Test.....	16
Step 2: Modify Testing Parameters (optional) .....	17
Step 3: Specify the Microbeam Properties (Occasional) .....	20
Step 4: Load the Specimen.....	21
Step 5: Set the Optics Zoom Level .....	22
Step 6: Adjust the Initial Size Crosshairs .....	22
Step 7: Select a Tracking Point.....	23
Step 8: Execute the Test.....	24
Step 9: Terminate the Test Prematurely (optional).....	24
6. Additional Settings .....	25
Camera Shutter and Gain .....	25
Grid Overlay .....	25
Temperature .....	26

Configuring the Real-Time Graphs .....	27
Data Output Settings .....	28
Advanced Settings.....	29
7. Reviewing Test Results .....	30
Overview .....	30
Selecting Images .....	31
Image Playback Options.....	31
Data Overlay.....	32
Exporting Images and Movies.....	33
Image Tracing.....	34
Exporting Traced Data .....	36
8. System Calibration and Advanced Tools.....	37
Piezo Calibration.....	37
Optics Calibration .....	38
Snap Image .....	38
9. Mounting a MicroBeam .....	40
10. Force Calculation Formulas .....	47
11. Appendix A: System Setup .....	48
Unpack Main Unit, Controllers, and Accessories.....	48
Foam Removal from Piezo Stack.....	49
Attach Spacer to Vertical Piezo Stage .....	50
Camera Unpacking and Assembly.....	51
Remove Lens Cap and Mount the Ring Light on the Lens Assembly .....	57
Set Test Chamber on the Chamber Holder .....	58
Place Temperature Sensor in the Test Chamber .....	58
Connect Cables .....	59
System Calibration Check.....	60
12. Appendix B: Software Installation.....	61
Update Firmware .....	68
13. Appendix C: Shear Testing .....	70
14. Appendix D: Camera Lens Attachment .....	74
15. Appendix E: Attaching Secondary Camera .....	79
16. Appendix F: Software Troubleshooting .....	83

# 1. General Information

Electrical Input to Power Supply	100-240VAC, 50-60Hz
Input Current Rating of Power Supply	1.0 Amp
Electrical Output from Power Supply	24 VDC
Output Current Rating of Power Supply	1.5 Amp
Recommended Computer OS	Windows 10
Environmental Conditions	Maximum Operating Temperature 25°C 20% - 80% Relative Humidity
Installation Category	Category II
Pollution Degree	Degree II
Data Connections	1 – USB 3.0 (SS) for camera/PC communication 1 – USB for Main system/PC communication USB hubs may be used for computers that have insufficient ports, as long as the hub supports and provides USB 3.0 connections.

## ***System Assembly***

Please see instructions in Appendix A for details on system assembly. See Appendix B for installing the software, and Appendix C for details on setting up the host computer.

## ***Connections to Supply***

Connect power cords to properly grounded 100-240VAC power source to ensure safe operation. Ensure that the power cords are always easily accessible. The use of an Uninterruptible Power Supply (UPS) is recommended to protect against data loss.

## ***Safety Warnings***

This equipment must be used in accordance with the procedures outlined in this manual to prevent injury and/or damage.

## ***Operating Controls***

There are no manual controls on the MicroTester G2 main unit. Both the camera and microbeam assembly contains motorized X, Y and Z adjustment stages to position and focus the camera view. In addition, the camera brightness can be controlled by setting the software interface.

## ***General Maintenance***

Clean the system as needed with mild soap and water or alcohol-based cleaning solutions.

## ***Approvals and Certification***

We declare that the CellScale MicroTester G2 complies with the requirements of the Electromagnetic Compatibility Directive, 2004/108/EC by demonstrating compliance with the following standards to the extent applicable:

EN 55011: 2008 Class A

EN 61000-3-3: 2008

EN 61000-4-2: 2009

EN 61000-4-3: 2006 (+A1:2008 +A2:2010)

EN 61000-4-4: 2005 (+A1: 2010)

EN 61000-4-5: 2006

EN 61000-4-6: 2009

EN 61000-4-8: 2010

EN 61000-4-11: 2004

## 2. Parts

### ***Tools***

Small flat head screwdriver

SAE hex wrenches: 5/64", 3/32"

Metric hex wrenches: 1.5mm, 2mm, 2.5mm, 3mm

Rectangular flat wrench

### ***Basic Components***

Main Assembly

Piezo Stage Controller

Position Sensor Controller (with LED power supply)

USB 3.0 Colour Camera

Motorized XYZ Camera stages and Controller

### 3. Testing Terminology

The MicroTester G2 is designed to apply compressive loads to small test specimens 50 to 2000  $\mu\text{m}$  in size. Typical specimens include biological tissue samples, scaffolds, cell aggregates, or hydrogel microspheres.

#### ***Multiphase Test Cycles***

In order to properly characterize and test a specimen, it is often necessary to load it to different degrees and at different rates. There are three main reasons for doing this:

***Preconditioning*** - The goal of preconditioning is to restore a specimen to its physiological or *in vivo* state. During the process of specimen storage and preparation, a specimen may swell, dry out, have its material fibers realign, or its molecules reorganize. It may take multiple preconditioning cycles for a specimen to be restored to its natural state.


***Reproducing Physiological Conditions During Testing*** - By applying various loads and load rates, natural expansion and contraction of a specimen can be reproduced (for example, the pulse pressure in an organ). In as much as the physiological conditions can be recreated, the specimen can be tested in a more realistic state.

***Varying Test Conditions*** - Variable loads and rates allow you to create a variety of test profiles to best study your specimen.



## Phases, Cycles, and Test Sequences

As the following diagram demonstrates, each application and release of load on the specimen is called a **test cycle**. The same test cycle can be repeated multiple times to achieve a certain goal (preconditioning, physiological conditioning, or testing); this is called a **test set**. Finally, a **test sequence** is made up of multiple test sets.



The above example describes the following:

- The entire diagram presents a full *test sequence*.
- Within that sequence, there are two *test sets*.
- Within the first set, two identical *test cycles* are implemented to bring the sample to a satisfactory in vivo state.
- Finally, the second set is made up of three cycles.

## Test Phases: The Smallest Unit of Testing


The test phase is the smallest unit of the test specification. There are four phases within a cycle. Each phase serves a specific purpose:

**Loading** - During the load phase, a deformation is applied to the specimen. The deformation can be specified either in terms of force applied or displacement achieved.

**Holding** - The deformation can be held for a given duration. The duration for which it is held is dependent on the nature of the testing.

**Recovering** - The recovery phase is the time during which the force being applied to the specimen is removed. The duration of the recovery time is configurable and dependent on the nature of the testing.

**Resting** - Finally, the rest phase is the time between the end of one cycle and the beginning of the next. Some tests may specify a short recover time, while others may specify a longer time. The duration is configurable and dependent on the nature of the testing.



## Control Modes

There are two control modes which define the basic approach to a given test: displacement control and force control:

Under **displacement control**, the displacement of the specimen is predefined. The MicroTester G2 compresses the specimen until the predefined displacement is achieved. The force required to achieve the displacement is an output of the test.


Under **force control**, the force applied to the specimen is predefined. The MicroTester G2 compresses the specimen until the predefined force is achieved. The displacement required to achieve the force is an output of the test.

## Control Functions

The MicroTester G2 makes it possible to test specimens under several control functions with Displacement or Force control modes:


The **ramp function** applies the displacement at a constant nominal rate. This is equivalent to a constant engineering or nominal strain rate.

The **step function** achieves and maintains the desired force as quickly as possible. The amount of time it takes to achieve the desired force depends on the material being tested and the control settings.



## 4. System Overview

The MicroTester G2 is a mechanical test device that can be used to test the properties of a variety of tissues and materials.



The main components of the device are the load application microbeam, the actuators, the imaging system, and the fluid bath/test chamber.

### **Microbeam**

The microbeam is rigidly fixed to the vertical actuator (which is under closed loop control). It has a flexible shaft that is typically between 30mm and 50mm long and can be made from a length of metal, glass, or plastic with a circular or rectangular cross section. Circular tungsten beams are supplied with the system. At the end of the shaft is a rigidly fixed compression plate. This plate is typically made from a small square of micro-cut stainless steel and is adhered to the end of the shaft. Cyanoacrylate and adhesive accelerator are provided for this purpose.

The microbeam is used to apply deformations to the test specimen while measuring the applied forces. The deformation is determined by measuring the difference between the current tip position and the initial tip position. These measurements are obtained in real time at a frequency of 5 Hz by image tracking techniques. The applied force is determined by calculating the deflection of the microbeam (the difference between the displacement of the tip and the displacement of the actuator). The displacement of the actuator is obtained directly from a position feedback device (displacement sensor) mounted to the actuator.

Because the microbeam is used for 2 purposes (apply deformation and measure forces), an iterative control scheme is necessary to achieve the desired result. For example, if a tip displacement of 100µm is desired, it is unknown at the start of the test how much actuator travel will be required to move the tip by this amount since some of the stage travel will result in tip displacement and some will result in microbeam bending.

The microbeam can have a variety of constructions and must be tailored to the force ranges that are to be measured. If the microbeam is too stiff, then the beam deflection will be too small. This has a negative influence on the force resolution. This is a result of the position resolution of the base and tip displacement

measurements, which are both approximately  $0.1\mu\text{m}$ . For example, if the microbeam only deflects  $2\mu\text{m}$  during the test, then the force resolution is only  $1/20$  of the maximum force.

Conversely, if the microbeam is too flexible, then the deflection will be too large. This can have a negative influence on the reaction time of the device and on the parallelism of the compression plates. With more deflection of the microbeam, the actuator must travel further to achieve the desired tip displacement. Since the velocity of the actuator is limited to about  $2\text{mm/s}$ , this can cause unwanted delays. Also, as the beam deflects, the top compression plate can become slightly less parallel to the lower compression plate. In a test with a moderate microbeam deflection of  $100\mu\text{m}$ , this effect is very small, since the beam is typically over  $30\text{mm}$  long. In a test with a large microbeam deflection of  $2\text{mm}$ , this effect becomes more significant.

The table below shows a list of the microbeams that come with the MicroTester G2 system and the appropriate maximum forces for each diameter. A chapter containing details of the force calculations can be found later in this manual.

Diameter (mm)	0.0762	0.1016	0.1524	0.2032	0.3048	0.4064	0.5588	1.0160	1.5748
Modulus (MPa)	411000	411000	411000	411000	411000	411000	411000	411000	411000
Typical Force Resolution ( $\mu\text{N}$ )	0.012	0.038	0.191	0.604	3.056	9.658	34.52	330.54	1907.88
Recommended Max Force ( $\mu\text{N}$ )	4	11	57	181	917	2,897	10,536	99,162	572,363



*Microbeam with Platen*

## ***Microbeam Actuators***


There are 2 actuation systems integrated to the microbeam assembly. The first system is comprised of the 2 piezo motor driven actuators that are used to position the microbeam in plane before the start of the test. The actuators are controlled by the user, using on-screen controls in an open loop jog mode. There is no absolute measurement of how far they have been moved since feedback is based on user interpretation of the live on-screen image.

The second actuator system is the closed-loop, piezo motor actuator that is used to move the microbeam vertically. It is driven in a user-controlled jog mode during setup and then under computer control during the test (specially modified systems are also available to track horizontal displacement as well, in which case the horizontal actuator is also under closed loop control). The actuator used for this purpose is identical to the other 2 axes, but the system also incorporates a position feedback device. This displacement sensor has 0.1 $\mu$ m resolution and allows for the software to determine the displacement of the actuator relative to the start of the test. This motion constitutes the movement of the base of the microbeam. The displacement of the base and tip of the microbeam are needed in order to determine the deflection of (and hence the force applied by) the microbeam.

## ***Imaging System***

The imaging system is comprised of a USB digital camera, a zoom lens, and a motorized XYZ stage assembly to allow position adjustment and focusing. The camera has a resolution of 2048 X 2048 pixels and can match the control frequency of 5Hz. The zoom optics allow for a field of view ranging from approximately 0.4mm to 12.0mm.

The imaging system comes with interchangeable tubes and auxiliary zoom lenses for 1X and 2X magnification. How to change between lenses can be found in Appendix E.




## **Fluid Bath**

The fluid bath contains approximately 45mL of fluid and is heated by a resistance foil heater in the base of the bath. The fluid temperature is controlled by an onboard temperature controller that communicates with the PC via the USB connection. The temperature of the bath is measured by a solid-state temperature probe with a temperature resolution of 0.1°C. To function the MicroTester G2 software will log the actual temperature of the fluid. To enable the controller to heat the fluid, the “HEATER” switch beside the main power switch must be turned on to illuminate the amber heater light.

### **MicroTester G2 Tip: Temperature Probe**

*The temperature controller works best if the temperature probe tip is completely submerged and **not** touching the side of the fluid chamber.*


The bath is a clear polycarbonate dish which allows for visibility from all sides, the main imaging system is setup to image the front side of the dish, where the compression platform is setup in the image below. The compression plate is typically made of a stainless-steel block and is flush with the front window.



The software included with the MicroTester G2 system is called MicroTester G2. It is divided into two modules: a data collection module and a review and analysis module. The data collection module is used to set test parameters, enable specimen loading and testing, and monitor test progress. The screen layout for this module is shown below:

## Test Parameters

## Live Output




## Live Video

## Live Graphs



The review and analysis modules are used to playback accelerated or decelerated test images, perform image analysis and tracking, and output movie files for presentation purposes. The screen layout for this module is shown below:



## Output Files and Data Structures

For each test, the MicroTester G2 creates and saves three file types. The following table describes the five file types for a project named “Sample1”. Output from this test would be found in a “Sample1” output directory (a sub directory of the user specified data directory).

File Type	Description
.tst file	The Sample1.tst file that contains the exact protocol and settings used for the test.
.csv file	The Sample1Data.csv file that contains comma separated numerical data such as time, force and displacement values.
.jpg files	Captured images such as Sample1.0001000.jpg, which would correspond to an image captured at 1000 milliseconds from the start of the test.

Each test folder will also contain 2 subfolders. The “Logs” subfolder contains a text file of the content of the text dialog portion of the main screen. It is useful as reference to troubleshoot problems with your system, should they occur.

The “Tracking” subfolder is initially empty. If tracking is done on any of the images in the test folder using the “Analyze and Review” software module, there will be data files that contain the tracking information stored in this directory.

While using the “Analyze and Review” software module, you may create additional data files such as text files (\*.csv) containing tracking results, images with force data or tracking results overlaid (\*.wmf), or movie files (\*.avi). These additional files can be stored in the test directory or elsewhere on your computer’s hard drive without interfering with the software application (once created, they cannot be opened by the MicroTester G2 application).

## 5. Setting Up & Starting a Test

### ***Overview***


Setting up and running a new test is a simple process made up of both mandatory and optional steps. The following list presents all the steps, while the rest of this chapter describes each of the steps in detail.

- Step 1: Start a New Test
- Step 2: Modify the Test Parameters (optional)
- Step 3: Specify the Microbeam properties (occasional)
- Step 4: Load the Specimen
- Step 5: Set the Optics Zoom Level
- Step 6: Adjust the Initial Size Crosshairs
- Step 7: Select a Tracking Point
- Step 8: Execute the Test
- Step 9: Terminate the Test Prematurely (optional)

## Step 1: Start a New Test

Launch the MicroTester G2 software, and then select Collect New from the File menu. In the Create Test From Template dialog, perform the following steps:

1. Select a template that matches the type of test you wish to perform. See the *MicroTester G2 Tip* on how to select and use a template. You can modify the template parameters in Step 2, below.
2. Name your test. The dialog will have a default test name. You can use the default or rename the test. Each time you start a new test, the default name will continue to increment the number at the end of the name.
3. If desired, you can change the location of the output data. The location of your output data and images is determined by specifying a Test Name and Data Directory. The template and data directories are user specific (computer login name). Each system user can store their files to a different default location. The system uses the last location selected as the default.



### **MicroTester G2 Tip: Selecting and Using Templates**

*How to select a template: Designing an appropriate test sequence is an art that depends on both the type of material being tested and the specific material properties you are interested in measuring. When first testing a new material, you should expect to have to experiment with the settings until the test yields meaningful data.*

*The system comes loaded with example templates to help you get started. Selecting a template does not lock you into a specific test sequence or protocol – rather a template defines a test sequence and settings, all of which can be changed before a test is run. Once you have developed a test sequence and settings that are appropriate for the material you are testing, you should save these settings using “Save As Template” from the file menu. You can then select your template the next time you initiate a test.*

## Step 2: Modify Testing Parameters (optional)

You can select and modify test parameter sets by clicking on their row in the Test Sequence table and then pressing the Edit Set button (or by double clicking on their row).

Test Sequence										Temperature Set Point	Total Time
<div>Edit Set   Delete   Insert Above   Insert Below</div>										37	00:00:45
Set Name	Axis	Magnitude	Load(S)	Hold(S)	Recover(S)	Rest(S)	Reps	Data(Hz)	Image(Hz)	Parameters	
Data1	Z-Compress	Ramp 200.0 (um)	20	5	20	0	1	5	1		

When you do so, the Set Parameter Editor Dialog will appear.

Note that displacements are specified in either % strain, or  $\mu\text{m}$ . Force loads are specified in  $\mu\text{N}$ , and durations are specified in seconds.

The following table describes each of the parameters.

### Set Parameter Editor

Name  (Example: "Squish1", "TestData1")

Axis

Control Axis

Control Mode

Control Function

Compress Magnitude

Tolerance (um)

☐ Correct Overshoot

Cycle

Loading Duration (S)  ☐ Do Not Load Repetitions

Hold Duration (S)  Data Output Frequency (Hz)

Recovery Duration (S)  ☐ Do Not Recover Image Output Frequency (Hz)

Rest Duration

OK

Cancel

Test Parameter	Description
<b>Control Axis</b>	Loading is normally applied in the vertical Z-Compress axis. MicroTester G2 systems that have been equipped with a second displacement control axis can also apply horizontal shear loads by selecting the X-Shear axis.
<b>Control Mode</b>	<p>Compression tests are typically performed under displacement control. However, you can select the following control modes to achieve specific testing objectives:</p> <ul style="list-style-type: none"> <li>• For creep testing, use force control with a long hold duration.</li> <li>• For relaxation testing, use displacement control with a long hold duration.</li> </ul>
<b>Control Function</b>	Ramp functions are most commonly used although step functions are also used.
<b>Compress Magnitude</b> (Load Magnitude when in Force Control mode)	<p>Compression magnitude is specified as an absolute position or force relative to the initial state of the microbeam. Positive magnitudes are compressive (the beam moves down). In systems equipped to perform horizontal shear testing, positive magnitudes are to the right.</p> <p>If you are using Displacement control mode, you can specify the displacement in either <math>\mu\text{m}</math> or as a percentage. For example, the displacement on a <math>1000\mu\text{m}</math> Initial Size specimen can be expressed as either <math>100\mu\text{m}</math> or 10%.</p> <p>If you are using Force control mode, you can only specify the force magnitude in <math>\mu\text{N}</math>.</p> <p>Watch the test results carefully to determine which magnitude setting best achieves your test goals.</p>
<b>Tolerance</b>	<p>For step functions only.</p> <p>A step compression is considered achieved when the current microbeam position (or force) is at the specified magnitude +/- the tolerance. A tolerance that is too small may result in unacceptably long compression durations and/or instability.</p> <p>Recommended values are at least 10 times the force resolution of the beam. (see chapter: Force Calculation Formulas for more details).</p>
<b>Correct Overshoot</b>	<p>For step functions only.</p> <p>An overshoot is defined as a compression magnitude that is larger than the specified magnitude plus the tolerance. When checked the control software will move the microbeam up and/or down until the desired step magnitude is achieved (i.e. recover from overshoot). When unchecked, the control software will consider overshoot as the successful completion of the step function.</p> <p>Not correcting overshoot is important for relaxation testing.</p>
<b>Loading Duration</b>	For evenly spaced images, it is recommended to choose a number that is an integer multiple of the Image Output Frequency.
<b>Hold Duration</b>	Hold Duration is typically set to 0, however it is useful for creep or relaxation testing.
<b>Recovery Duration</b>	Recovery Duration is often set to the same value as the Stretch Duration.
<b>Do Not Recover</b>	Select this to skip the recovery phase.
<b>Rest Duration</b>	Rest duration is typically set to 0, however a non-zero value may be used to mimic <i>in vivo</i> conditions or for specialized testing.
<b>Repetitions</b>	Apply enough repetitions until the force deformation curves from one repetition to the next start to overlap each other.
<b>Data Output Frequency</b>	Typically set to the same frequency as the image output frequency.
<b>Image Output Frequency</b>	Typically set to 1Hz for cycles > 30 seconds and 5 Hz for cycles < 30 seconds.

**MicroTester G2 Tip: Multi-Stage Compression Testing**

Recover phases can be skipped by checking the “Do Not Recover” check box. For example, a set that performed 10% compression with the “Do Not Recover” box checked would result in the specimen still being compressed by 10% at the start of the next cycle.

Multi-Stage compression can be achieved by performing multiple sets that have a “Do Not Recover” duration.

For example, to compress a specimen by 300um in 3 stages of 100um, create 3 separate sets (i.e. three separate lines in the test parameter specification table). The first set should have a 100um compression magnitude, the second a 200um compression magnitude, and the third a 300um compression magnitude. (Note the compression magnitude is relative to the start of the test rather than the position at the start of the set).

**MicroTester G2 Tip: Testing Tissues that Relax Quickly**

For specimens that relax very quickly, it may be necessary to achieve the desired move in 2 or 3 stages so as not to overshoot and rebound. What can happen in these cases is that the microbeam will achieve its desired displacement, completing the “Compress” phase. At the start of the hold phase, the tendency of the specimen is to relax, causing a drop-in force. This dropping force causes the microbeam to want to unload its elastic energy and hence further deform the specimen. The software will react to this by raising the stage (microbeam base) to reduce the applied force and maintain the desired displacement. The difficulty occurs when the relaxation causes these events to happen so quickly that the 5Hz operation of the equipment cannot keep up and maintain the desired displacement. The only remedy is to move the beam part way to its destination in 1 or more preliminary moves so that the overshoot that takes place never results in the microbeam moving more than its desired displacement.

**MicroTester G2 Tip: Applying Preloads**

Preloading can be implemented by creating a set with a force control step function load at the beginning of a testing sequence.

**MicroTester G2 Tip: Tension Testing**

The MicroTester G2 can be used to perform tension testing by fixing the specimen to the lower and upper compression plates. To achieve this, the compression magnitude should be entered as a negative number.

### Step 3: Specify the Microbeam Properties (Occasional)

Microbeam properties including dimensions and modulus must be correctly specified to obtain correct force outputs. You can check and set Beam properties by selecting Equipment from the Settings menu or by pressing the Equipment Settings button in the Live Output section of the main display window. The following dialog will appear.

Note that the “Beam Length” parameter refers to the distance from the fixed end of the microbeam wire (the near side of the clamp) to the point of force application at the free end under the compression platen.

The microbeam length described above **MUST** be accurately measured. Calipers or a finely marked ruler can be used.

Standard microbeam diameters and moduli are shown in the table below. Potentially, other microbeams of other materials could be used. Microbeams may need to be cut before mounting to between 30-60 mm in length. Scissors or side-cutters may be used for this purpose.

**For instructions on preparing and mounting a microbeam, see Section 9 of the manual - Mounting Microbeams.**

The Equipment Settings dialog box contains the following sections and controls:

- Beam Properties:** Radio buttons for Rectangular and Circular (selected). Input fields for Length (mm) = 58, Thickness (mm) = 0, Diameter (mm) = 0.5588, and Modulus (MPa) = 411000.
- Camera:** Shutter (mS) = 5, Gain (0-100) = 10, Resolution = 2048x2048 24 bit color. Checkboxes for Auto are present for Shutter and Gain. A note states: Recommended Maximum Shutter = 10 mS.
- Auxiliary Camera:** Shutter (mS) = 0, Gain (0-100) = 0, Resolution = (empty dropdown).
- Temperature:** Setpoint (°C) = 37. A Warnings section includes a checkbox for Show Warnings if Temperature Not within Setpoint Tolerance (unchecked) and a Tolerance (Degrees C) = 1.
- Grid Overlay:** A checkbox for Show (unchecked) and a spacing (pix) = 100.
- Optics:** Pixels Per um Per Zoom = 0.33.
- Buttons for OK and Cancel at the bottom.

Diameter (mm)	0.0762	0.1016	0.1524	0.2032	0.3048	0.4064	0.5588
Modulus (MPa)	411000	411000	411000	411000	411000	411000	411000
Typical Force Resolution (μN)	0.012	0.038	0.191	0.604	3.056	9.658	34.521
Recommended Max Force (μN)	4	11	57	181	917	2,897	10,536


#### **MicroTester G2 Tip: Update Templates After Changing Beam Properties**

*It is important to update the beam properties after replacing a microbeam. **The beam properties are stored in the test template along with the test protocol and other equipment settings.** For these updated beam properties to be automatically used the next time you start a new test using the same template, you should save these settings to a new template using “Save As Template” from the File menu.*



## Step 4: Load the Specimen

Loading the specimen under the microbeam takes a certain amount of practice. It is important to position the specimen as close as possible to the front glass so that it can be imaged correctly. The end of the microbeam can be used as a micromanipulator for this purpose. The microbeam can be moved using the manual Actuator Controls which allow the user to move each actuator independently at a variety of jog speeds.



### **MicroTester G2 Tip: Viewing the Specimen in the Live Video Window**

*The optics and camera are mounted on three axis manual stages which can be used to focus and position the optics and camera to obtain optimal imaging.*

*The optics have a manual zoom control detent which can be used to change the magnification of the image and increase the field of view. It is often helpful to work at low zoom levels for loading a specimen. The zoom level can then be increased once the specimen is properly loaded.*


*Camera Shutter and Gain can also be adjusted (on the Equipment Settings dialog) if the live image appears too bright or too dark. See Section 6 for more details.*

### **MicroTester G2 Tip: Microbeam Position**

*For most testing it is desirable to move the microbeam forward until it touches the glass and then move it backward a very small amount. It is usually obvious when the microbeam touches the glass because it will jump suddenly in the live image.*

## Step 5: Set the Optics Zoom Level

A zoom specification grid is located below the Actuator Control interface. The “Zoom” grid **MUST** be set correctly (at the same value as the zoom indicator on the lens) so that the software can convert image tracking data from pixel values to  $\mu\text{m}$  values.



### **MicroTester G2 Tip: Setting a Default Software Zoom Level**

*The initial software zoom level is stored in the test templates. This value will be saved to the current selected value when saving a template using “Save As Template” from the File menu.*


## Step 6: Adjust the Initial Size Crosshairs

There are two red crosshairs overlaid on the live image. If the zoom level is set correctly, the “Initial Size” field on the left will display the distance between these crosshairs. The crosshairs can be moved on the screen by clicking on them and dragging them (they will turn yellow when they are clickable and can be dragged).

Usually the crosshairs are moved to be positioned at the top and bottom of the test specimen.

Note that the absolute distance is measured between the crosshairs, not the linear distance in one axis. Thus, the crosshairs must be aligned horizontally if a correct vertical distance is to be measured.

Initial Size is one of the outputs in the .csv data output file.




## Step 7: Select a Tracking Point

Select an appropriate tracking point on the microbeam or compression plate by **right-clicking** on the live video window. A green rectangle or square will appear showing the region that will be used to track this location. The selected point should have some contrast (light/dark) within the green rectangle and should not change in appearance due to lighting glare (i.e. it should be a real feature, not an effect created by the lighting). Since force calculations are based on the tracked point locations, accurate tracking is essential for accurate force outputs.


Good choices for the tracking point location, as shown in the following figure, include locations on the microbeam wire end or on the compression plate attached to the microbeam. Features in focus tend to track better than features that are out of focus or blurry.

Good choices on the microbeam include regions of high contrast within the rectangle. A poor location (shown in red) on the microbeam does not have enough contrast within the rectangle for the tracking algorithm to uniquely determine its location.

Good choices on the compression plate will span across a sharp boundary with, for example, the bottom half of the green tracking rectangle covering the compression plate and the top half of the tracking rectangle not on the compression plate. Poor choices on the compression plate (shown in red) include locations that are overlapping with or close to the specimen (which may change in appearance during compression) or do not span across one of the plate edges.




## Step 8: Execute the Test

Select Execute from the File menu, or click  on the toolbar.

### **MicroTester G2 Tip: Check Settings**

*Before executing a test, be sure that the microbeam properties and optics zoom level are set correctly. Accurate results are highly dependent on these settings. Also check the position of the initial size crosshairs and tracking point.*

## Step 9: Terminate the Test Prematurely (optional)


You can stop a test at any time by clicking  on the toolbar, or by selecting Stop from the File menu.

## 6. Additional Settings

### Camera Shutter and Gain

Camera settings can be adjusted by selecting Equipment from the Settings menu or by pressing the Equipment Settings button in the Live Output section of the main display window. The following dialog will appear.

Increasing the camera shutter duration or the camera gain makes the image brighter. Decreasing either of these parameters makes the image darker. If the gain is too large, the image may appear grainy. If the shutter is too long, the image may appear blurry. These settings are stored with the test template file.




The Equipment Settings dialog box is divided into several sections. The 'Beam Properties' section at the top has radio buttons for 'Rectangular' and 'Circular', with 'Circular' selected. Below are input fields for 'Length (mm)' (58), 'Thickness (mm)' (0), 'Diameter (mm)' (0.5588), and 'Modulus (MPa)' (411000). The 'Camera' section includes 'Shutter (mS)' (5), 'Gain (0-100)' (10), and 'Resolution' (2048x2048 24 bit color). It also has 'Auto' checkboxes for shutter and gain, and a note: 'Recommended Maximum Shutter = 10 mS'. The 'Auxiliary Camera' section has similar controls for 'Shutter (mS)' (0), 'Gain (0-100)' (0), and 'Resolution'. The 'Temperature' section has a 'Setpoint (°C)' (37) and a 'Warnings' box with a checkbox for 'Show Warnings if Temperature Not within Setpoint Tolerance' and a 'Tolerance (Degrees C)' (1). The 'Grid Overlay' section has a 'Show' checkbox and a 'spacing (pix)' (100). The 'Optics' section has a 'Pixels Per um Per Zoom' (0.33). At the bottom are 'OK' and 'Cancel' buttons.

### Grid Overlay

Within the Equipment Settings menu, a grid overlay option can be implemented during a trial. The spacing can be manually inputted (in pixels). Checking the Show check box will display a green grid with the specified spacing.

Applying a grid overlay can provide many benefits during a trial such as distance referencing, beam alignment checking, and quadrant mapping.



## Temperature

To activate the fluid chamber heater, not only must the ON/OFF switch be turned on, but the “ENABLE” button must be pressed so that the HEATER ON indicator is illuminated. At this point, the heater will warm the fluid to the last used temperature set point. The temperature set point can be changed by selecting Equipment from the Settings menu.

Typically, the heater will take 10 to 30 minutes to warm a full fluid chamber from room temperature to 37°C.

When Show Temperature Warnings is enabled, and the Current Temperature is not at the Temperature Set Point plus or minus the specified tolerance -- a red indicator button will be displayed next to the Current Temperature live output and a warning dialog will appear when you execute the test.

Test Sequence

Edit Set

Delete

Insert Above

Insert Below

Temperature Set Point

37

Total Time

00:00:45

Set Name	Axis	Magnitude	Load(S)	Hold(S)	Recover(S)	Rest(S)	Reps	Data(Hz)	Image(Hz)	Parameters
Data1	Z-Compress	Ramp 200.0 (um)	20	5	20	0	1	5	1	

## Configuring the Real-Time Graphs

The three graphs on the right side of the screen provide real-time user feedback during a test. These graphs are intended to be used for qualitative feedback, not for detailed analysis.


The graph settings can be specified by selecting Graphs from the Settings menu. Auto scaling allows the axis min and max to start at the specified values but expand if the data values are larger or smaller than the initial limits.

The image shows a 'Graph Settings' dialog box with a green title bar and a close button. It contains three sections for 'Graph 1', 'Graph 2', and 'Graph 3'. Each section has an 'XAxis' and a 'YAxis' configuration area. The 'XAxis' area includes a dropdown menu, an 'Auto' checkbox, and 'Min' and 'Max' input fields. The 'YAxis' area includes a dropdown menu, an 'Auto' checkbox, and 'Min' and 'Max' input fields. Below each section is a 'Clear Graph Between Sets' checkbox. At the bottom of the dialog are 'OK' and 'Cancel' buttons.

Graph	XAxis Variable	XAxis Min	XAxis Max	YAxis Variable	YAxis Min	YAxis Max	Clear Graph Between Sets
Graph 1	Time (S)	0	40	Force (mN)	0	500	<input type="checkbox"/>
Graph 2	Time (S)	0	40	Displacement (um)	0	400	<input type="checkbox"/>
Graph 3	Displacement (um)	0	400	Force (mN)	0	500	<input checked="" type="checkbox"/>

It is typical to display Force versus Time, Displacement versus Time and Force versus Displacement as shown below. However, any of the following variables can be plotted in any of the three graphs:

Force ( $\mu$ N)	Displacement ( $\mu$ m)	Time (S)
Force (mN)	Displacement (mm)	Temperature



## Data Output Settings

Select Data Output from the Settings Menu and the following dialog will appear.

Users can specify the \*.jpg image file quality that will be used for saving collected images.

**Data Output Settings**

**Format**

☐ tab delimited (.xls)

☒ comma delimited (.csv)

**Output Columns**

**Images**

JPEG Quality  (0-100)



## Advanced Settings

Select Advanced from the Settings menu and the following dialog will appear.

Hold Phase Control allows the user to set some limits on actuator movement, in order to improve system stability.

The first parameter limits the maximum movement of the vertical stage during each control loop iteration.

The second parameter ("Slow Ratio") is like a safety factor. For example, if a move of 10 $\mu$ m is desired in each time step, the actuator will attempt to move 10 $\mu$ m X the slow ratio. See the tip below for more information.

Image Tracking Settings allow users to control the size of the search region and the size of the tracking template. These parameters should normally not be changed.

For systems equipped to measure the X-Shear axis, a square template and search region is recommended.

Advanced Settings

Hold Phase Control

☐ Limit Moves To a Maximum of 50 Steps Per 200 mS Interval

Limiting moves may cause target values to be reached more slowly  
But can be set to ensure that each move completes within each 200 mS Interval  
Allowing for full 5Hz data output with no skipped points

0.75 Slow Ratio

Corrections for overshoot or undershoot are multiplied by the Slow Ratio  
If this ratio is small target values may be achieved slowly  
If this ratio is large oscillations about the target value may occur

Image Tracking Settings

Search Region Minimum (+/- pixels) 3 x 30

Template Region Size (+/- pixels) 5 x 21

Increasing these settings will increase real time computations during data collection  
And may result in 5Hz cycle time not being achieved

OK Cancel

### MicroTester G2 Tip: Slow Ratio

*For the hold phase of linear elastic materials, both the beam tip and beam base will remain close to stationary during a hold phase, so a slow ratio of 0.3-0.5 is appropriate to improve stability. For highly viscoelastic materials, the beam tip will tend to keep moving downward at the end of the compression phase, resulting in increasing compression and decreasing applied force (due to decreasing beam deflection). To counteract this, the software will have to move the beam base upward for a displacement-controlled test (to maintain the current displacement) or downward for a force-controlled test (to maintain the current force). In these situations, a higher slow ratio may be needed. A ratio of 0.75-1.0 is relatively stable. Higher ratios will be even more responsive, but also potentially more unstable. Trial and error may be required to obtain the optimal results for your material.*

# 7. Reviewing Test Results

## Overview

The MicroTester G2 software has an integrated image analysis module which can be accessed by selecting Analyze and Review Images from the File menu, then selecting the appropriate test file. The test file is a text file with a .tst extension that contains information about the test parameters so that the software can display the images and data in a useful fashion.

The “Analyze and Review Images” module allows for post processing of the test images. It is used for 3 main purposes:


1. Review test for quality assurance purposes
2. Generate movies and other material for presentations
3. Trace specimen profiles to calculate geometry values

Image List

Image Playback

Image Tracing

Data Overlay



## Selecting Images

The Images are shown in a tree structure on the top left panel. The tree structure organizes the images by:

**Set** (line on the test parameter window)


**Cycle** (iteration of a given Set)

**Phase** (Compress/Hold/Recovery/Rest of each Cycle)

**Image** (the one or more individual images from each Phase)


You can select which images to include in a playback set in one of several ways:

- Hold down the shift key and click on individual images
- Select entire sets, cycles, or phases (will playback all contained images)
- Select images which have data associated with a specific tracking set



## Image Playback Options


The controls on the left of the screen are broken into 2 sections: Playback and Tracking. The playback buttons allow the user to review the images taken during a test.




Next/Prev	Display next or previous image (as dictated by "Play Every" parameter).
Play/Stop	Start and stop the playback.
Prev/Next Cycle	Jumps to a corresponding image and phase in the next or previous cycle. Useful when reviewing sets that have multiple cycles.
Loop	Sequences can be played continuously in a loop or only once.
Playback Rate	+/- changes the speed of the playback. A playback rate can also be manually entered in the display box.
Play Every	Allows the users to skip some frames to expedite playback.

## Data Overlay

Data associated with each image is displayed below the tracking controls on the left side of the main window. Clicking the Show Graph check box will display the collected force data associated with this test as a graph overlaid on the images. Clicking on the Show Tracking Point check box will display the tracked point used during the collection of this test. Since force calculations are based on the tracked point locations, a test where the tracked point suddenly jumps or loses its correct location indicates invalid force outputs.

Clicking the details icon  displays the dialog shown below. The user can specify the graph's size, location and appearance. The data points associated with the current image are marked on the graph with green circles. The overlaid data graph will be drawn on top of any image tracking displays (when they overlap). This can be useful for correlating force events with image events and for generating presentation material.



Data

0.00

ZForce (uN) XForce (uN)

Compress Shear

☐ Show Graph

☐ Show Tracking Point

Data Output Settings

Data

Force (uN)

Graph Overlay

Range

☐ Current Set

☐ Current Cycle

☒ Entire Test

Size (Pixels)

Width 400

Height 100

Location (Pixels)

Left 400

Top 400

ZCompress

☒ Show

Line Width (pixels)

2 Color

XShear

☒ Show

Line Width (pixels)

2 Color

☐ Show Solid Background Color

☒ Show Border Color


OK Apply Cancel

## Exporting Images and Movies

The current Image view, including all displayed tracing data (see section on Image Tracing) and overlaid data can be exported as a metafile (wmf) by selecting Current Image from the Export menu. Alternately, the current image view can be copied to the clipboard by selecting Ctrl-C from the keyboard.

To export a series of image views, including all displayed tracing data and overlaid data as an .avi movie, select Movie from the Export menu. The dialog shown on the right will be displayed. The image list is automatically set to the currently selected images in the test image list tree. This image list can be changed by selecting new images in the test image list tree and pressing the set button. Since there can be many images in a single set or phase, clicking the “Reduce” button can be useful to remove every other frame in the Image List. This button can be used multiple times to pare down the image list so that the movie file will not be too large.

Options are available for setting the frame rate or image size (resolution), adding filenames, and choosing color or grayscale output. Compression options are also available.



The "Create Avi Video" dialog box is used for exporting image views as an .avi movie. It features an "Image List" on the left with a list of image files (0000000.jpg to 0010000.jpg) and buttons for "Set", "Add", "Delete", "Clear", and "Reduce (Remove Every Other Frame)". The "Appearance" section on the right includes a "Frame Rate (fps)" field set to 10, an "Image Size" dropdown set to 50%, and checkboxes for "Add Image Titles (Filenames)", "Grayscale" (selected), and "Color". The "Compression" section has a "Compress" checkbox and an "Advanced" button. The "Output Filename" field at the bottom shows "D:\EquipmentTesting\squish\Data\Test958\Movie1.avi" with a "Browse" button. "OK" and "Cancel" buttons are at the bottom.



Section	Field/Control	Value/Option
Image List	Image List	0000000.jpg 0001000.jpg 0002000.jpg 0003000.jpg 0004000.jpg 0005000.jpg 0006000.jpg 0007000.jpg 0008000.jpg 0009000.jpg 0010000.jpg
	Buttons	Set, Add, Delete, Clear, Reduce (Remove Every Other Frame)
	Frame Rate (fps)	10
	Image Size	50%
	Add Image Titles (Filenames)	<input type="checkbox"/>
Appearance	Grayscale	<input checked="" type="radio"/>
	Color	<input type="radio"/>
Compression	Compress	<input type="checkbox"/>
	Advanced	Advanced
Output Filename	Output Filename	D:\EquipmentTesting\squish\Data\Test958\Movie1.avi
Buttons	OK	OK
	Cancel	Cancel

## Image Tracing





Image Tracing is a function that can be used to determine aggregate geometries and surface tensions during cell aggregate compression testing.

The tracing portion of the screen allows the user to create traces, which are associated with a specific image. If the “Display” checkbox is checked, then any traces will show up on the screen when the images are being viewed. The “Select Images” button will automatically select all of images that have associated traces. This is useful if your data set is large and the traced images are spread out.

Clicking “Create” (or if an existing trace exists, clicking “Edit”) brings up the following dialog. This dialog floats so that it can be positioned to allow the user to visualize the image and the dialog simultaneously.




The general procedure for creating a specimen tracing is as follows:

<p>1. While the “Add” button is depressed, select several points around the perimeter of the image.</p>	
<p>2. Click the “Close Curve” button to attach the first and last point.</p>	
<p>3. Use the “Move” and/or “Insert” buttons to modify the points as needed. There is an image zoom feature available on the top toolbar if needed.</p>	
<p>4. Press “Calculate” to generate the curve fit geometries.</p>	
<p>5. Use the line location adjustment arrows to position the vertical lines such that they intersect the circles at the point where the aggregate surface intersects the compression plates.</p>	
<p>6. Click “OK”</p>	

## Exporting Traced Data

Geometries and Surface Tensions calculated using the image tracing module can be exported to a .csv text file by selecting Tracing Results from the Export menu.

The default list contains the images that were selected when the “Tracing Results” menu item was selected. Items can be added or deleted using the dialog buttons.





## 8. System Calibration and Advanced Tools


### *Piezo Calibration*

Piezo Calibration can be selected from the Calibrations submenu of the Tools menu.

While the vertical piezo stage relies on the displacement sensor for determining the amount of travel, a calibration routine is necessary so that the software can estimate the correlation between movement pulses and displacement. This calibration routine will attempt to move the piezo actuator a certain amount and will measure the actual movement achieved. It is recommended that the piezo calibration be performed when the equipment is being set up for the first time in your facility. It is also required to be performed whenever the MicroTester G2 software is used on a new computer as the calibration factor is stored on the computer.

Before calibration, make sure the vertical piezo stage is in the mid-range of its travel, preferably at the height from which a test will be started.


The calibration results are reported in the log window on the left side of the main screen.



## Optics Calibration

Optics calibration must be completed when the system is first set up and every time the camera and lens system is disassembled or changed. It must also be performed every time a new computer is used to run the MicroTester G2. The procedure is as follows:


1. Launch the MicroTester G2 software and launch a new test in order to see the live video window.
2. Bring up the Optics Calibration Dialog by selecting Optics Calibration from the Calibrations submenu of the Tools menu.
3. Position a calibrated target with features of known size on the front of the fluid chamber. A glass target with 10 lines/mm (1mm = 1000  $\mu\text{m}$ ) is provided with your equipment for this purpose. Note that this means 10 lines/mm. This target can be positioned with provided holder. Focus the Image using the three axis stages on which the optics and camera are mounted.
4. Drag the crosshairs on the live video window such that they are positioned a known distance apart. Be sure that the crosshairs are **aligned with each other** as the computer measures the **absolute distance** between them, not just the horizontal or vertical distance.
5. Enter the distance between the crosshairs in the dialog field.
6. Select the optics zoom level.
7. Click the “Calibrate” button in the dialog box.
8. Selecting the “Apply” button will cause the software to use the current calibration value. It is useful to compare the current calculated value with the previous value as the value should not change dramatically. A typical calibration value is around **0.32 pixels/ $\mu\text{m}$ /zoom**.



## Snap Image

Single Images can be captured and saved from the live video stream by selecting Snap Image from the Tools menu.

This feature allows the user to capture a single image during the setup phase. This can be useful for documenting specimen loading activity and neutral (non-contact) beam position.



## 9. Mounting a MicroBeam

The exact details of how to best adhere the upper compression plate and mount the microbeam to the grips may vary depending on your specific application, but a general protocol is outlined here:


1. Gather the following materials:
  - a. Vertical mounting stage
  - b. 1.5mm Hex key
  - c. 5/64" hex key
  - d. Cyanoacrylate (i.e. Crazy Glue or Loctite 422)
  - e. Glue Accelerator
  - f. Cotton Swab
  - g. Tweezers
  - h. Microbeam clamp with M3 fastener
  - i. Desired microbeam and compression plate



2. Bring the vertical stage to its highest point by rotating the knob on the left side. If the stage does not move, release the stopper on the right side by rotating and loosening the side screw.




3. Align the microbeam clamp to the lip on the face of the stage, and fasten it using a 2.5mm hex key and M3 fastener



4. Loosen the side screw of the microbeam clamp using a 5/64" hex key




5. Place microbeam below the clamp opening and lower the stage until the beam is just above the tips of the clamp. Tighten the clamp using a 5/64" hex key. This is to ensure the microbeam is horizontal. Lift the stage slightly to ensure it is secure affix in the clamp




***MicroTester G2 Tip: Microbeam Position***

*Very thin and/or delicate microbeams will tend to sag under their own weight (e.g. 100 $\mu$ m diameter and smaller metal microbeams). To ensure that the upper compression plate is level after mounting, it may be desirable to initially place the microbeam on an inclined surface.*

6. Place the desired compression plate below the tip of the microbeam. Use tweezers to help ensure the tip of the microbeam is in line with the edge of the compression plate




7. Place a small dab of glue on the tip of a cotton swap, and place this on the tip of the microbeam



**MicroTester G2 Tip: Cyanoacrylate Adhesive**


*Cyanoacrylate adhesive is provided complimentary with the MicroTester G2 system to bond the microbeam and compression platen. This adhesive will best retain its properties if it is **refrigerated** when not being used.*

8. Lower the microbeam until it touches the plate
9. Cure the glue by brushing a small amount of glue accelerator





10. Lift the stage and remove the clamp from the stage using a 2.5mm hex key to loosen the M3 fastener



11. The clamp and microbeam assembly is ready to be mounted to the MicroTester system using the M3 fastener and 2.5mm hex key


***MicroTester G2 Warning: Adhesive and Accelerator***

*Cyanoacrylate adhesive and accelerator are provided with the MicroTester G2 system to bond the microbeam and compression platen. These are eye, skin, and respiratory irritants. Cyanoacrylate will instantly bond skin. The use of gloves is recommended.*




The microbeam clamp is attached to the vertical piezo stage with a single M3 fastener. This single connection allows for the angle of the upper compression platen to be adjusted under guidance from the live image. It is important that the upper compression platen and the lower compression anvil are parallel for proper compression testing.


The microbeam length must be measured and the dimensions and modulus correctly specified before a test. See Section 3 - Step 5 for more details.



*Clamp with microbeam and platen attached to the vertical piezo stage, with testing anvil and sample*



*Clamp with MicroBeam and platen attached to the vertical piezo stage, with testing anvil and sample*



*Live software camera view MicroBeam and platen, with testing anvil and sample*

## 10. Force Calculation Formulas

The force applied to the specimen is calculated using the deflection of the microbeam according to the formula:

$$Deflection = \frac{Force \times Beam Length^3}{3 \times Beam Modulus \times Beam Area Moment of Inertia}$$

where

$$Beam Area Moment of Inertia (round) = \frac{\pi \times radius^4}{4}$$

$$Beam Area Moment of Inertia (rectangular) = \frac{Width \times Height^3}{12}$$

Typically, the microbeams are made from tungsten wire. CellScale has this wire available in 7 different diameters. With a beam length of 55mm, the expected force resolutions and recommended maximum forces are given below.

Diameter (mm)	0.0762	0.1016	0.1524	0.2032	0.3048	0.4064	0.5588	1.0160	1.5748
Modulus (MPa)	411000	411000	411000	411000	411000	411000	411000	411000	411000
Typical Force Resolution (μN)	0.012	0.038	0.191	0.604	3.056	9.658	34.52	330.54	1907.88
Recommended Max Force (μN)	4	11	57	181	917	2,897	10,536	99,162	572,363

# 11. Appendix A: System Setup


## *Unpack Main Unit, Controllers, and Accessories*



Remove the main unit from its packing material. A foam “clamshell” encases the main unit inside a cardboard box for shipping. This material must be carefully removed. Take care not to lift the main unit by the camera lens or positioning knobs.

The temperature controller, piezo stage controller, control box, and all system cables and accessories are packed in a separate container.

## ***Foam Removal from Piezo Stack***



The piezo actuator stack and displacement sensor are protected by a plastic cap. Foam surrounds these components inside the cap during shipping. Remove the cap using a hex key and remove all pieces of foam inside. Replace the cap and fasteners.




### ***Attach Spacer to Vertical Piezo Stage***




A spacer is attached to the vertical piezo stage. This allows the microbeam clamp to reach the test chamber. 4 x M3x6 fasteners are used to attach this component using the supplied 2.5mm hex key.

## ***Camera Unpacking and Assembly***


Remove the plastic cover over the camera stage using the long T-handle 3mm hex key to loosen 3x M4 screws.




Remove the foam surrounding the camera stage and inside the camera cover.



The stage is held in place with zip ties and plastic spacer blocks for shipping. Cut the zip ties shown below.




First slide the Left/Right stage in the direction below to remove the plastic spacer block. **Gently retract the stages back towards the motor.** The stages are spring loaded and releasing them at full tension can result in damage to the motors and/or linear stage itself.





Next slide the Forward/Backward stage in the direction below to remove the plastic spacer block. **Gently retract the stages back towards the motor.** The stages are spring loaded and releasing them at full tension can result in damage to the motors and/or linear stage itself.




Unpackage the camera assembly and attach to camera mount with the included M4 fasteners.



The camera assembly can now be fastened on to the stages using an 3mm hex key.




The camera USB cable should now be attached and tightened down with the thumb screws.



Run USB cable through back hole of top cover. Start lowering the cover unto the camera assembly with the ring LED connector placed through the front hole of cover.




Next attach the camera cover using the long 3mm hex key provided fasten the 3 x M4 cover fasteners located below.






## ***Remove Lens Cap and Mount the Ring Light on the Lens Assembly***

Remove the lens cap from the lens assembly. Attach the ring light to the end of the lens assembly using the 3 set screws spaced around the ring light.




### ***Set Test Chamber on the Chamber Holder***



Set the fluid bath (test chamber) in the chamber holder on the main unit. This chamber is pre-assembled with a glass slide and gasket. Replacements are provided for these components.

### ***Place Temperature Sensor in the Test Chamber***

Bend the temperature sensor into the test chamber. This chamber can be filled with fluid for which temperature control can be specified. The temperature sensor functions best when it is fully immersed in fluid and is not touching the side of the chamber.



## Connect Cables

Few power and communication cables need to be connected between the main unit, a power supply, and a computer:

- 20V power supply to power input jack on the main unit
- LED ring light to control box power extension
- USB Cable to back of camera (do not yet connect to PC)
- USB Cable for Piezo board, Arduino DUE board, and position sensor (control box hub)


NOTE: At the completion of this step, there should be 2 USB connections ready to connect to the PC USB ports

- Black USB cable from camera (this will require a USB 3.0 port)
- Black USB cable from USB hub in control box controlling displacement sensor module, Arduino DUE board, and Piezo control board.



Camera USB

Remember to tighten camera USB connection by rotating the 2 pin fasteners on the cable.



## ***System Calibration Check***


After the MicroTester G2 is fully assembled and set up, the MicroTester G2 software must be installed (see Appendix B). The system can then be switched on and the computer connectivity may be tested by opening an initial test. However, before any tests are run, it is recommended to check the piezo and optics calibrations. Instructions for these operations can be found in Section 8: System Calibration and Advanced Tools.





## 12. Appendix B: Software Installation

1. Download the software installation package from the provided link and save to a known location
2. Right click on the .zip installation file, right click, and select “Extract All...” to unzip the installation package
3. Go to the uncompressed folder and select “setup.exe”.
4. Follow the prompts as shown on the next few pages.
5. Once the installer is finished, plug the USB cables into a USB port for the following components:
  - Camera (black USB cable)
  - Position Sensor, Piezo Actuator Controller, and Arduino DUE USB hub cable

The MicroTester G2 requires at least 2 USB ports for operation.





MicroTester - InstallShield Wizard

**Customer Information**

Please enter your information.

User Name:  
MicroTester

Organization:  
Cellscale

InstallShield

< Back Next > Cancel

MicroTester - InstallShield Wizard

**Ready to Install the Program**

The wizard is ready to begin installation.

If you want to review or change any of your installation settings, click Back. Click Cancel to exit the wizard.

Current Settings:

Setup Type:  
Typical

Destination Folder:  
C:\Cellscale\MicroTester\

User Information:  
Name: MicroTester  
Company: Cellscale

InstallShield

< Back Install Cancel

# Device Driver Installation Wizard



## Welcome to the Device Driver Installation Wizard!

This wizard helps you install the software drivers that some computers devices need in order to work.


To continue, click Next.

< Back

Next >

Cancel

Depending on security settings, the following message may appear.



# Device Driver Installation Wizard



## Completing the Device Driver Installation Wizard




The drivers were successfully installed on this computer! If a device came with your software, you can now connect it to this computer. If your device came with instructions, please read them first.




Driver Name	Status
✓ Newport CDM Driver Pa...	Ready to use
✓ Newport CDM Driver Pa...	Ready to use




< Back

Finish

Cancel








## Update Firmware

Periodic firmware updates may be issued by CellScale for the MicroTester. When a new firmware or software is available, you may download it from our website ([www.cellscale.com/downloads](http://www.cellscale.com/downloads)).

To update the firmware:

1. Connect the MicroTester to the PC via USB and turn on power to the MicroTester
2. Open the *UpdateFirmware.exe* program located in *MicroTester>Software>Utilities>Update Firmware*
3. The program will automatically update the firmware of the MicroTester
4. Power cycle the system after the firmware is successfully updated before using the MicroTester software again

The firmware version is checked and the automatic updater is also run when installing a new version of the software.





If the automatic firmware update software does not work, please follow the manual steps below:

1. With the MicroTester connected to the PC and turned on, launch the manual firmware update software located in *MicroTester>Utilities>Update Firmware>Advanced*.
2. Click the *Connect* button.
3. Load the firmware file using the *Browse* button. The file will be a .bin extension.
4. Execute the update using the *Update* button.
5. Power cycle the system after the firmware is successfully updated before using the MicroTester software again.

The screenshot shows a software window titled "UpdateFirmware" with a close button (X) in the top right corner. Below the title bar, it displays "Version 1.5" and "Copyright (C) 2009-2021". The interface is divided into three main sections, each with a numbered heading and a button:

- 1. Connect To Device**: Contains a "Connect" button and an empty text input field.
- 2. Select Firmware File**: Contains a "Browse" button and an empty text input field. A small green dot is visible to the right of this section.
- 3. Update Firmware**: Contains an "Update" button and an empty text input field. Below the input field is a large, empty rectangular area with scrollbars, likely for displaying progress or logs.


At the bottom of the window, there is a "NOTE: device must be power cycled after updating firmware" and a "Done" button.

## 13. Appendix C: Shear Testing



Some MicroTester G2 systems are equipped with an additional displacement sensor to measure horizontal displacement of the microbeam during a test so that tests can be executed in this axis. This allows a type of shear testing to be performed on samples.

On systems equipped for shear testing, a special version of the software is provided directly to the user (the link on the CellScale website contains the single-axis version of the software). In the “Set Parameter Editor”, this allows X-Shear to be selected as the Control Axis. Compression and Shear movements can be executed in the same test sequence by selected a different Control Axis for different test sets. For shear testing, positive magnitudes are to the right.



The image shows a 'Set Parameter Editor' dialog box with a close button (X) in the top right corner. The dialog is divided into two main sections: 'Axis' and 'Cycle'.

**Axis Section:**


- Name:** A text field containing 'SHEAR'. To its right is a hint: '(Example: "Squish1", "TestData1")'.
- Control Axis:** A dropdown menu with 'X-Shear' selected.
- Control Mode:** A dropdown menu with 'Displacement' selected.
- Control Function:** A dropdown menu with 'Ramp' selected.
- Shear Magnitude:** A text field with '100' and a unit dropdown menu with 'um' selected.
- Tolerance (um):** A text field with '5'.
- Correct Overshoot:** A checked checkbox.

**Cycle Section:**


- Loading Duration (S):** A text field with '30'. To its right is an unchecked checkbox labeled 'Do Not Load'.
- Hold Duration (S):** A text field with '0'.
- Recovery Duration (S):** A text field with '10'. To its right is an unchecked checkbox labeled 'Do Not Recover'.
- Rest Duration:** A text field with '0'.
- Repetitions:** A text field with '1'.
- Data Output Frequency (Hz):** A dropdown menu with '1' selected.
- Image Output Frequency (Hz):** A dropdown menu with '1' selected.

At the bottom of the dialog are two buttons: 'OK' and 'Cancel'.

Additional data columns for the X-Shear axis are output in the data .csv file for these specially equipped systems. In addition, the live graphs during a test will contain additional data output for this axis.



Also, the image tracking settings will typically differ from that of a single axis MicroTester G2 system. The tracking template and search region settings can be found in Advanced in the Settings menu. This will be preset as a square dialog box under the demo template provided so that image tracking can be performed equally well in both axes. If single axis testing is being performed, these settings can be altered so that the template and search region is optimized for the axis being used.



The image shows a software dialog box titled "Advanced Settings" with a close button (X) in the top right corner. The dialog is divided into two main sections: "Hold Phase Control" and "Image Tracking Settings".

**Hold Phase Control**

- ☐ Limit Moves To a Maximum of  Steps Per 200 mS  
Limiting moves may cause target values to be reached more slowly  
But can be set to ensure that each move completes within each 200 mS  
Allowing for full 5Hz data output with no skipped points
- Slow Ratio  
Corrections for overshoot or undershoot are multiplied by the Slow Ratio  
If this ratio is small target values may be achieved slowly  
If this ratio is large oscillations about the target value may occur

**Image Tracking Settings**

- Search Region Minimum (+/- pixels)**  
 x
- Template Region Size (+/- pixels)**  
 x

Increasing these settings will increase real time computations during data  
And may result in 5Hz cycle time not being achieved.

At the bottom of the dialog are two buttons: "OK" and "Cancel".

## 14. Appendix D: Camera Lens Attachment

The MicroTester G2 system features a camera lens tube with an operating detent mechanism. This prevents accidental slippage during experimentation, giving a consistent and crisp focus view of the intended target. The standard MicroTester G2 comes with an already-attached camera assembly. This standard assembly includes the camera bracket (short), camera, objective tube (1x), and auxiliary lens (1x). The standard assembly is depicted below.



The standard MicroTester G2 also comes with the 2x camera objective tube, 2x mounting bracket, and 2x auxiliary lens for different applications seen below. Changing either the objective tube OR the auxiliary lens will have the effect of doubling the magnification. Change BOTH will have the effect of increasing the magnification by a factor of 4.





- a. In order to replace the auxiliary lens, slowly rotate the attached piece counter clockwise. Be careful not to touch any exposed lens. Attach the wanted auxiliary lens by slowly rotating the part clockwise onto the camera lens tube until a snug fit is apparent. Changing the auxiliary lens will change the working distance of the optics and thus require changing the mounting bracket.





- b. In order to replace the objective tube, first loosen the 2 set screws (circled in red) from the camera bracket and take off camera assembly from MicroTester G2 device.




Remove the camera from the objective tube by slowly rotating (counter clockwise) the camera.



To change between 1x and 2x objective tubes:


1. Locate ring B which only has 2 screws and loosen them
  - a. **Do not loosen screws on ring A and C!**



2. Gently pull the tube assembly out and replace it with the desired tube



- a. Ensure the new tube is pushed up against the side of ring A



3. Tighten both screws at B
4. Screw the lens assembly back onto the camera
5. Place the assembly into the appropriate bracket and tighten the set screws to hold it in place
  - a. Use the longer bracket when using the 2x objective tube, and the shorter bracket with the 1x tube


Take camera system and line detent clip with the horizontal set screw. Carefully push camera system back into the intended bracket. The short bracket is meant for the 1x auxiliary lens while the longer bracket is for the 2x auxiliary lens. Rotate camera system counter-clockwise until detent clip is facing downwards. Tighten set screws.



## 15. Appendix E: Attaching Secondary Camera

An additional option for the MicroTester G2 is a secondary camera axis that allows you to view your sample from the side. This option includes an anodized aluminum bracket what mounts to the side of the device. This bracket is help on by 4x M4x12 fasteners.

To attach this bracket the bottom cover must first be removed, this is done by carefully laying the system on its side. It is recommended to support the piezo motor cover with a piece of foam or something soft, so you do not scratch the surface. Once the system is on its side remove 6 M3x6 flat head fasteners using the provided 2mm hex key




The bracket will attach to the 4 holes shown below.






When the system is on its side set the bracket in place and fasten it with provided fasteners.




Attach the camera bracket to the top of the XYZ stage using the supplied M4 fasteners.



Slide the additional spacer over the lens. Place the camera lens through the camera bracket and secure it in place by tightening the M3x8 set screws with a 2.5mm hex key. Ensure the spacer is place behind the bracket as shown below.

Secure the secondary ring light over the lens.



Connect the USB cable to the camera and plug it into a USB 3.0 port on your computer.

To utilize the secondary camera feature, start a new test with the 'demo2Camera' template provided with the MicroTester software. You may edit this template as needed to suit your sample and testing requirements. Remember to save your changes by using the 'Save as a new Template' under the 'File' drop down menu.

Create Test From Template ×

Select

demo2Camera

Test Name

Test\_001

Data Directory


C:\CellScale\MicroTester\Data Browse

Template Directory

E:\Dropbox (CellScale)\Manufacturing Info\MicroSquisher\ Browse

OK Cancel

If you are running a test near the edge of the media dish, it may be difficult to focus the secondary camera due to the radius of the corner on the plastic dish. We recommend sliding the dish forward slightly to allow the secondary camera to focus on your specimen.




Depending on where your MicroTester is set up, it may be helpful to place a small piece of white paper along the left edge of the dish (when looking at the beam head on). This will eliminate any background objects from being captured and help with reflecting light towards the camera.



## 16. Appendix F: Software Troubleshooting

This software is recommended for use with a Windows 10 operating system. Use on Windows 7 systems may result in some hardware communication errors with the MicroTester, such as the following:

**Issue:** MicroTester software cannot communicate with camera and/or controller



**Solution:** This is often caused by turning off the device while the software is still running on Windows 7, then trying to reopen the software. If you encounter this error, try closing and turning everything off, removing and reinserting the USBs, then try again. Make sure both USBs and the power supply are plugged in and the device is on before you launch the software. You can also try manually renumbering the port numbers for the USB port being used by the MicroTester controller in your computer's Device Manager > Ports. Finally, try restarting your computer.